

MRI Guidelines

Enterra® Therapy

Gastric Electrical Stimulation System

Patient Guide



Rx only

Note: This document contains information related to Magnetic Resonance Imaging (MRI) use with the Enterra® Therapy Gastric Electrical Stimulation (GES) System. Refer to the product manuals for more detailed information on non-MRI aspects of implantation, programming, charging, and use of the components of the Enterra System.

enterra
medical™

Glossary



MR Conditional - An item with demonstrated safety in the MR environment within defined conditions, including conditions of the static magnetic field, the switched gradient magnetic field and the radio frequency fields. Additional conditions, including specific configurations of the item, may be required.






MR Unsafe - An item which poses unacceptable risks to the patient, medical staff, or other persons within the MR environment.

ACR Zone - Zones of an MR site that denote areas with various MR safety levels, as defined by the American College of Radiology.

MRI - Magnetic Resonance Imaging.

Table of Contents

GLOSSARY.....	2
1. WHAT IS MAGNETIC RESONANCE IMAGING (MRI)?	4
2. CAN I HAVE AN MRI?	4
3. MRI SAFETY INFORMATION 	5
3.1. MR conditional devices 	6
3.1.2. For MRI examinations using a detachable head, upper extremity, or lower extremity RF transmit/ receive volume coil.....	6
3.2. MR unsafe devices 	6
4. POSSIBLE RISKS OF MRI WITH THE ENTERRA THERAPY SYSTEM.....	9
4.1. Heating of the implant	9
4.2. Unintended stimulation.....	10
4.3. Image distortion and artifacts	10
4.4. Magnetic field interactions.....	10
4.5. Device malfunction or damage	11
5. MRI GUIDELINES.....	11
5.1. Before starting MRI scan.....	11
5.2. During MRI scan	13
5.3. After MRI scan	13

1. WHAT IS MAGNETIC RESONANCE IMAGING (MRI)?

Magnetic Resonance Imaging (MRI) is a technique that is used for creating pictures of the internal structures of the body. Unlike an X-ray exam, it does not use radiation. Instead, it uses a large magnet, radio waves, and a computer to create pictures of body structures and organs.

2. CAN I HAVE AN MRI?

Patients with an implanted Enterra® Therapy System may have an MRI scan of some body parts under certain conditions. Consult your doctor to determine if you are eligible for MRI examination.

You are required to discuss with your doctor if you have any other device(s) implanted. Possible implanted devices include:

- Pacemaker or Implantable Cardioverter-Defibrillator (ICD)
- Some aneurysm clips
- Cochlear implants
- Orthopedic prostheses (e.g., hip implant)

- Other neurostimulators
- Stents
- Metal plates, pins, or screws
- Dental implants

An MRI requires the patient to lie still during the exam. Before the MRI procedure, inform the MRI technologist of the following:

- If you are pregnant or suspect you are pregnant
- If you are breast feeding at the time of the scheduled procedure
- If you are having a fever

3. MRI SAFETY INFORMATION

The Enterra System is **MR Conditional**. This means that patients with the Enterra System can safely have MRI examinations of some body parts under certain conditions. The conditions for MRI scans will vary with the type of MRI coil.

Always obtain the latest MRI guidelines. Refer to the contact information on the last page of this manual or go to www.enterramedical.com/hcp/manuals.

3.1. MR Conditional Devices

Non-clinical testing has demonstrated that the Enterra II 37800 Neurostimulator with two Enterra Therapy 4351 35 cm unipolar leads are **MR Conditional**.

Patients with these devices can be safely scanned in an MR system meeting the conditions in “For MRI Examinations Using a Detachable Head, Upper Extremity, or Lower Extremity RF Transmit/Receive Volume Coil”.

3.1.2. For MRI examinations using a detachable head, upper extremity, or lower extremity RF transmit/receive volume coil

A patient implanted with the Enterra Therapy System may be safely scanned at the head or upper/lower extremity at 1.5T MRI under the following conditions. Failure to follow these conditions may result in injury to the patient.

3.2. MR unsafe devices

The external components of the Enterra Therapy System, including the Model 8840 Clinician Programmer, are **MR Unsafe**. These devices must NOT be taken into the MR scanner room (ACR Zone IV).



MRI Safety Information

A person with the Enterra Medical Enterra II System may be safely scanned under the following conditions. Failure to follow these conditions may result in injury.

Device Name	Enterra II
Device Configuration	Stimulation OFF
Static Magnetic Field Strength (Bo)	1.5 T
MR Scanner Type	Cylindrical
Bo Field Orientation	Horizontal
Maximum Spatial Field Gradient	20 T/m (2,000 gauss/cm)
Maximum Gradient Slew Rate	200 T/m/s per axis
RF Excitation	Circularly Polarized (CP)
RF Transmit Coil Type	Head or Extremity Coils Only
RF Receive Coil Type	Any
Operating Mode	Normal Operating Mode
Maximum Whole-Body SAR	2 W/kg
Maximum Head SAR	3.2 W/kg
Scan Duration	Scan for up to 60 minutes of continuous RF (a sequence or back to back series/scan without breaks).
Scan Regions	Only head or extremity scanning allowed.
MR Image Artifact	The presence of this implant may produce an image artifact.

Please consult your doctor and the MRI technologist to make sure that the specific conditions above are met before MRI examination.




Figure: MRI scan* using detachable **1** head, **2** upper extremity, or **3** lower extremity RF transmit coil.

*Illustrated in Figure are typical use scenarios of detachable RF transmit/receive volume coil. Other scanning scenarios are also permissible according to MR scanner/coil manuals. For example, an MRI scan of the ankle with a detachable lower extremity RF transmit/receive volume coil or an MRI scan of the upper arm with a detachable lower extremity RF transmit/receive volume coil in the superman posture is permissible given the aforementioned scan conditions are met.

4. POSSIBLE RISKS OF MRI WITH THE ENTERRA THERAPY SYSTEM

Non-clinical testing has shown that patients with the Enterra Therapy System can safely have MRI when the conditions for safe MRI described in this document are followed. However, there may be some risks of performing MRI when you have an implanted Enterra Therapy System. Possible risks include:

- Heating of the implant
- Unintended stimulation
- Image distortion and artifacts
- Magnetic field interactions
- Device malfunction or damage

4.1. Heating of the implant

MRI may cause the implant to become hot. However, if the conditions for safe MRI are followed, this heating is minimal. If the specific MRI conditions are not followed, heating of the implant could damage surrounding tissue. If the site of your implant feels hot during MRI, inform the MRI technologist immediately and then contact your doctor.

4.2. Unintended stimulation

MRI may cause unintended stimulation from the implant. This unintended stimulation may be uncomfortable (you may feel a tingling, shocking, or jolting sensation). However, if the conditions for safe MRI are followed, such stimulations may not happen. If you feel any uncomfortable stimulations during MRI, inform the MRI technologist immediately and then contact your doctor.

4.3. Image distortion and artifacts

No image distortion or artifacts should be seen from an MRI head and extremity scans.

4.4. Magnetic field interactions

The magnets used in MRI may cause the neurostimulator to shift or move slightly within the implant pocket. This may cause stress to tissues and/or the lead. As a result, you may feel a slight tugging sensation at the site of your implant. If you feel uncomfortable while in the MRI, inform the MRI technologist immediately.

4.5. Device malfunction or damage

Tests in various MRI systems were conducted. These tests did not cause any damage to, or malfunction of, the implant. If the implant malfunctions or becomes damaged, it may result in nerve damage and other associated problems. If you feel any stimulation or discomfort during MRI, inform the MRI technologist immediately and then contact your doctor.

5. MRI GUIDELINES

The guidelines for MRI scans are based on non-clinical tests conducted on the Enterra Therapy System.

Precautions are to be taken before, during, and after MRI scan. Talk to your MRI technologist or your doctor should you have any questions or concerns.

5.1. Before starting MRI scan

- Consult your doctor and MRI technologist to determine if you are eligible for MRI scan.
- Inform your doctor and MRI technologist if you have other medical device(s) implanted, such as a pacemaker, drug pump, hip prosthesis, stent, etc.

- Inform your doctor and MRI technologist if you think you have any the following conditions with your device: a broken lead fragment, lead disconnection from the neurostimulator, a partially implanted lead, a malfunctioning neurostimulator, a neurostimulator implanted at an area other than abdomen, or a system with open or low impedances (indicating a short circuit) on any electrodes. Consult your doctor for MRI eligibility if any of these conditions apply.
- Your MRI technologist may also give you MRI Patient Guides and Instructions. Make sure that you fully comply with the guidelines and instructions. Discuss any concerns with your MRI technologist or your doctor.
- Bring the most up-to-date patient ID card to all MRI appointments.
- For MRI using detachable head, upper extremity, or lower extremity RF transmit/receive volume coils, make sure that the neurostimulator stimulation is turned OFF.
- Make sure you remove any external metallic objects before entering the MRI room.

5.2. During MRI scan

- You may feel slight tugging, vibration, warming, and/or tapping in the area where the neurostimulator is located during the MRI scan. If those feelings cause discomfort, inform the MRI technologist know immediately.
- If you are not feeling well for any other reasons prior to or at the time of MRI scanning, please inform your MRI technologist.

5.3. After MRI scan

- After the MRI scan, have the stimulation turned back on.
- If you feel any changes in stimulation after an MRI and are uncomfortable, contact your doctor.

Notes



Manufacturer

Enterra Medical, Inc.
5353 W. Wayzata Blvd., #400
St. Louis Park, MN 55416
USA

www.enterramedical.com
Tel. +855-7-nterra or
+855-768-3772

Enterra® is a registered trademark of Enterra Medical, Inc. in the US, EU, and other regions.

©2023 Enterra Medical, Inc. All rights reserved.
800-00939-001, Rev B 2023-10

